


Overview EQA I/2019

Sample 2019-01: CT negative / NG positive




	RLU Panther
LABCON-OWL*	737
RefLab**	684
105	770
111	686
114	604
115	578
116	614
117	583
118	569
119	871
120†	1456†
122	637
123	763/1386†
124	647
125	659
128	807
130	796
131	733
133	683
134	661
135	854/621†
136	808
137	796

*RLU value constantly monitored over the entire testing phase of the external quality assessment by LABCON-OWL GmbH

**External Reference Laboratory in Germany

†Result for Aptima single tests of CT/NG, not included in the diagram

Sample 2019-02: CT negative / NG positive




	RLU Panther
LABCON-OWL*	1276
RefLab**	1273
105	1383
111	1207
114	1208
115	1292
116	1339
117	1269
118	1285
119	1418
120†	5706†
122	1134
123	1310/4722†
124	1283
125	1252
128	1433
130	1371
131	1289
133	1313
134	1338
135	1409/3585†
136	1330
137	1304

*RLU value constantly monitored over the entire testing phase of the external quality assessment by LABCON-OWL

**External Reference Laboratory in Germany

†Result for Aptima single tests of CT/NG, not included in the diagram

Sample 2019-03: CT positive / NG positive




	RLU Panther
LABCON-OWL*	2417
RefLab**	2530
105	2573
111	2434
114	2265
115	2373
116	2378
117	2407
118	2349
119	2591
120†	7324†/4547†
122	2119
123	2456/3629†
124	2464
125	2541
128	2616
130	2604
131	2517
133	2507
134	2603
135	2683/2492†
136	2518
137	2257

*RLU value constantly monitored over the entire testing phase of the external quality assessment by LABCON-OWL

**External Reference Laboratory in Germany

†Result for Aptima single tests of CT/NG, not included in the diagram

Sample 2019-04: CT positive / NG negative




	RLU Panther
LABCON-OWL*	1281
RefLab**	1344
105	1262
111	1360
114	1188
115	1190
116	1240
117	1259
118	1233
119	1356
120†	7133†
122	1152
123	1258
124	1378
125	1353
128	1355
130	1314
131	1355
133	1330
134	1392
135	1367
136	1290
137	1256

*RLU value constantly monitored over the entire testing phase of the external quality assessment by LABCON-OWL

**External Reference Laboratory in Germany

†Result for Aptima single tests of CT/NG, not included in the diagram


Appendix: Stability of the samples over the period of the external quality assessment



RLU values determined over the period of the external quality assessment by LABCON-OWL

The target value of all samples was confirmed and monitored in a reference laboratory in Germany before distribution. During the testing period the samples were retested three times by LABCON-OWL. The data are illustrated above. This graph shows the course of signal change for all four panel members over a period of nine weeks (storage at room temperature).

EQA Pilot for Mycoplasma genitalium



	RLU Panther
LABCON-OWL*	1881043
RefLab**#	Ct: 37 #
105	-
111	-
114	1672562
115	-
116	-
117	-
118	-
119	1963123
120	-
122	-
123	2196086
124	-
125#	Ct: 37,67 #
128	2014306
130	-
131	-
133	-
134	-
135	1807524
136	1940959
137	-

*RLU value constantly monitored over the entire testing phase of the external quality assessment by LABCON-OWL

**External Reference Laboratory in Germany

other than Hologic test kit, not included in the diagram