


Overview EQA II/2019

Sample 2019-05: CT negative / NG negativ




	RLU Panther
LABCON-OWL*	8
RefLab**	0
105	0
111	10
114	0
115	8
116	8
117	9
118	8
120†	2†
122	5
123	9
124	7
125	0
128	9
129	8
130	12
131	9
133	0
134	6
135	7
136	6
137	0
138	15
139	9

*RLU value constantly monitored over the entire testing phase of the external quality assessment by LABCON-OWL GmbH

**External Reference Laboratory in Germany

†Result for Aptima single tests of CT/NG, not included in the diagram

Sample 2019-06: CT positive / NG negative




	RLU Panther
LABCON-OWL*	723
RefLab**	860
105	940
111	1026
114	1033
115	818
116	765
117	973
118	819
120†	7386†
122	679
123	1023
124	805
125	744
128	926
129	592
130	912
131	452
133	556
134	889
135	1187
136	972
137	901
138	1157
139	844

*RLU value constantly monitored over the entire testing phase of the external quality assessment by LABCON-OWL

**External Reference Laboratory in Germany

†Result for Aptima single tests of CT/NG, not included in the diagram

Sample 2019-07: CT positive / NG positive




	RLU Panther
LABCON-OWL*	2696
RefLab**	2585
105	2416
111	2588
114	2523
115	2536
116	2442
117	2570
118	2511
120†	7520†/4897†
122	2507
123	2717/5387†
124	2707
125	2688
128	2744
129	2562
130	2511
131	2654
133	2594
134	2846
135	2798/6186†
136	2800
137	2452
138	2768
139	2445

*RLU value constantly monitored over the entire testing phase of the external quality assessment by LABCON-OWL

**External Reference Laboratory in Germany

†Result for Aptima single tests of CT/NG, not included in the diagram

Sample 2019-08: CT negative / NG positive




	RLU Panther
LABCON-OWL*	1248
RefLab**	1192
105	1078
111	1266
114	1198
115	1305
116	1285
117	1278
118	1245
120†	4043†
122	1133
123	1298/4437†
124	1248
125	1255
128	1227
129	1324
130	1169
131	1295
133	1193
134	1312
135	1267/5345†
136	1265
137	1237
138	1343
139	1153

*RLU value constantly monitored over the entire testing phase of the external quality assessment by LABCON-OWL

**External Reference Laboratory in Germany

†Result for Aptima single tests of CT/NG, not included in the diagram

Appendix: Stability of the samples over the period of the external quality assessment



RLU values determined over the period of the external quality assessment by LABCON-OWL

The target value of all samples was confirmed and monitored in a reference laboratory in Germany before distribution. During the testing period the samples were retested three times by LABCON-OWL. The data are illustrated above. This graph shows the course of signal change for all four panel members over a period of nine weeks (storage at room temperature).

